

Quality Control Ranges

MILLIPLEX® Human Fibrosis Panel 1

Catalog Number HFIB1-100K

QC Catalog Number/ Lot Number HFIB1-6100

<u>Hormone</u>	<u>QC Level</u>	<u>Lot #</u>	<u>Expected Range</u>	<u>Units</u>
IL-18BPa	Control I	HFIB1-104	70 - 145	pg/mL
	Control II	HFIB1-204	338 - 703	pg/mL
COMP	Control I	HFIB1-104	268 - 557	pg/mL
	Control II	HFIB1-204	1352 - 2809	pg/mL
Uteroglobin	Control I	HFIB1-104	141 - 292	ng/mL
	Control II	HFIB1-204	718 - 1492	ng/mL

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.

The vibrant M, Milliplex, Millipore and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates.

© 2026 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

Cat#: HFIB1-6100, Lot#: HFIB1-104.204 (23 FEB 2026)

